

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081822\
 Data File : VN073993.D
 Acq On : 18 Aug 2022 10:19
 Operator : JC\MD
 Sample : VN0818MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0818MBL01

Quant Time: Aug 18 21:46:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	629783	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	927943	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	888856	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	486719	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	305812	49.855	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.700%
35) Dibromofluoromethane	7.951	113	299998	54.667	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.340%
50) Toluene-d8	10.380	98	967947	45.543	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.080%
62) 4-Bromofluorobenzene	12.668	95	409955	58.119	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	116.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081822\
Data File : VN073993.D
Acq On : 18 Aug 2022 10:19
Operator : JC\MD
Sample : VN0818MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0818MBL01

Quant Time: Aug 18 21:46:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

