

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081822\
 Data File : VN074002.D
 Acq On : 18 Aug 2022 14:09
 Operator : JC\MD
 Sample : N4204-28DL 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-6DL

Quant Time: Aug 18 23:53:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	629854	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	922817	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	868661	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	452503	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	301703	49.179	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.360%
35) Dibromofluoromethane	7.951	113	295992	54.236	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.480%
50) Toluene-d8	10.380	98	966528	45.728	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.460%
62) 4-Bromofluorobenzene	12.669	95	398910	56.868	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	113.740%
Target Compounds						
11) Tert butyl alcohol	5.287	59	99424	72.453	ug/l	98
19) Methyl tert-butyl Ether	5.534	73	654890	36.579	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081822\
Data File : VN074002.D
Acq On : 18 Aug 2022 14:09
Operator : JC\MD
Sample : N4204-28DL 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-6DL

Quant Time: Aug 18 23:53:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

