

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081822\
 Data File : VN074005.D
 Acq On : 18 Aug 2022 15:23
 Operator : JC\MD
 Sample : N4241-07DL 25X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRE-1630DL

Quant Time: Aug 18 23:54:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	625419	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	932534	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	885428	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	487261	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	308969	50.721	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 101.440%	
35) Dibromofluoromethane	7.951	113	296208	53.710	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.420%	
50) Toluene-d8	10.380	98	969218	45.378	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 90.760%	
62) 4-Bromofluorobenzene	12.674	95	422339	59.580	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 119.160%	
Target Compounds						
					Qvalue	
16) Acetone	4.222	43	375456	130.642	ug/l	98
37) Ethyl Acetate	7.339	43	27320	2.944	ug/l	96
52) Toluene	10.445	92	5747	0.347	ug/l	84
68) m/p-Xylenes	11.892	106	4452	0.350	ug/l	95
84) 1,2,4-Trimethylbenzene	13.310	105	14940	0.518	ug/l	76
95) Naphthalene	15.439	128	14733	0.432	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081822\
Data File : VN074005.D
Acq On : 18 Aug 2022 15:23
Operator : JC\MD
Sample : N4241-07DL 25X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRE-1630DL

Quant Time: Aug 18 23:54:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

