

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081823\
 Data File : VN078939.D
 Acq On : 18 Aug 2023 13:47
 Operator : JC\MD
 Sample : VN0818MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0818MBL01

Quant Time: Aug 18 23:40:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 16 02:42:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	200494	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	363027	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	327584	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	100887	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	154335	49.469	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.940%
35) Dibromofluoromethane	8.171	113	130199	53.279	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.560%
50) Toluene-d8	10.571	98	442020	48.714	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.420%
62) 4-Bromofluorobenzene	12.853	95	130350	43.951	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081823\
Data File : VN078939.D
Acq On : 18 Aug 2023 13:47
Operator : JC\MD
Sample : VN0818MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0818MBL01

Quant Time: Aug 18 23:40:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 16 02:42:02 2023
Response via : Initial Calibration

