

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081823\
 Data File : VN078943.D
 Acq On : 18 Aug 2023 15:34
 Operator : JC\MD
 Sample : PB154902TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB154902TB

Quant Time: Aug 18 23:42:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 16 02:42:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	150567	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	276796	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	274905	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	82436	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	114690	48.951	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.900%
35) Dibromofluoromethane	8.171	113	97128	52.128	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.260%
50) Toluene-d8	10.571	98	371340	53.674	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.340%
62) 4-Bromofluorobenzene	12.853	95	105935	46.846	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.700%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	32197	33.305	ug/l #	81
20) Methylene Chloride	5.271	84	18622	9.039	ug/l #	79
25) 2-Butanone	7.494	43	9674	6.242	ug/l #	62
37) Ethyl Acetate	7.577	43	4876	1.555	ug/l #	67
43) Isopropyl Acetate	8.694	43	173304	31.267	ug/l #	79
51) 4-Methyl-2-Pentanone	10.453	43	44720	13.603	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081823\
Data File : VN078943.D
Acq On : 18 Aug 2023 15:34
Operator : JC\MD
Sample : PB154902TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB154902TB

Quant Time: Aug 18 23:42:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 16 02:42:02 2023
Response via : Initial Calibration

