

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081823\
 Data File : VN078948.D
 Acq On : 18 Aug 2023 17:35
 Operator : JC\MD
 Sample : PB154902ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB154902ZHE#05

Quant Time: Aug 18 23:45:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 16 02:42:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	146973	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	280216	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	276667	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	82965	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	118418	51.778	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 103.560%		
35) Dibromofluoromethane	8.171	113	97197	51.529	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.060%		
50) Toluene-d8	10.571	98	358691	51.213	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 102.420%		
62) 4-Bromofluorobenzene	12.853	95	105485	46.078	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 92.160%		
Target Compounds						
						Qvalue
16) Acetone	4.447	43	30149	31.949	ug/l	90
20) Methylene Chloride	5.277	84	17523	8.714	ug/l #	82
25) 2-Butanone	7.488	43	9211	6.088	ug/l #	83
37) Ethyl Acetate	7.577	43	4881	1.537	ug/l #	74
43) Isopropyl Acetate	8.694	43	209797	37.389	ug/l #	74
51) 4-Methyl-2-Pentanone	10.447	43	42099	12.650	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081823\
Data File : VN078948.D
Acq On : 18 Aug 2023 17:35
Operator : JC\MD
Sample : PB154902ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB154902ZHE#05

Quant Time: Aug 18 23:45:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081523W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 16 02:42:02 2023
Response via : Initial Calibration

