

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081825\
 Data File : VN087601.D
 Acq On : 18 Aug 2025 12:42
 Operator : JC\MD
 Sample : Q2845-04DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 002 35th Ave (july)DL

Quant Time: Aug 18 15:12:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 06 02:01:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.794	128	55447	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.082	114	303389	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.847	117	270310	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.559	65	153196	29.955	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	99.867%
60) 4-Bromofluorobenzene	12.829	95	128376	27.597	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	92.000%
63) Toluene-d8	10.547	98	375129	30.161	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.533%
Target Compounds						
					Qvalue	
62) Toluene	10.612	91	592568	35.518	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081825\
Data File : VN087601.D
Acq On : 18 Aug 2025 12:42
Operator : JC\MD
Sample : Q2845-04DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
002 35th Ave (july)DL

Quant Time: Aug 18 15:12:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 06 02:01:21 2025
Response via : Initial Calibration

