

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081919\
 Data File : VN057392.D
 Acq On : 19 Aug 2019 11:53
 Operator : JC/SP
 Sample : K4272-13
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2S-K-(1)

Quant Time: Aug 20 07:54:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 11:33:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1080558	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1492378	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1261074	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	320721	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	488271	41.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.14%	
35) Dibromofluoromethane	7.58	113	450670	46.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.14%	
50) Toluene-d8	10.09	98	1738167	48.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.16%	
62) 4-Bromofluorobenzene	12.40	95	461453	36.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.98%	
Target Compounds						
16) Acetone	3.81	43	30592	7.507	ug/l	93
20) Methylene Chloride	4.55	84	25468	2.234	ug/l	90
43) Isopropyl Acetate	8.16	43	235733	11.065	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081919\
Data File : VN057392.D
Acq On : 19 Aug 2019 11:53
Operator : JC/SP
Sample : K4272-13
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2S-K-(1)

Quant Time: Aug 20 07:54:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 11:33:14 2019
Response via : Initial Calibration

