

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081919\
 Data File : VN057394.D
 Acq On : 19 Aug 2019 12:42
 Operator : JC/SP
 Sample : K4339-15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T8-A2-(0)

Quant Time: Aug 20 07:59:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 11:33:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1095590	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1529955	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	1277085	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	340152	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	500411	42.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.04%	
35) Dibromofluoromethane	7.58	113	459389	45.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.62%	
50) Toluene-d8	10.08	98	1751775	47.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.52%	
62) 4-Bromofluorobenzene	12.40	95	471787	36.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.78%	
Target Compounds						
16) Acetone	3.81	43	17452	4.224	ug/l	100
20) Methylene Chloride	4.55	84	23635	2.045	ug/l	94
43) Isopropyl Acetate	8.16	43	231873	10.570	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081919\
Data File : VN057394.D
Acq On : 19 Aug 2019 12:42
Operator : JC/SP
Sample : K4339-15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T8-A2-(0)

Quant Time: Aug 20 07:59:26 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 11:33:14 2019
Response via : Initial Calibration

