

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081919\
 Data File : VN057400.D
 Acq On : 19 Aug 2019 15:07
 Operator : JC/SP
 Sample : K4339-21
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T8-A4-(4)

Quant Time: Aug 20 08:18:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 11:33:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1082121	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1510402	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1279450	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	342054	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	498264	42.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.72%	
35) Dibromofluoromethane	7.58	113	449798	45.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.86%	
50) Toluene-d8	10.09	98	1744952	48.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.38%	
62) 4-Bromofluorobenzene	12.40	95	466931	36.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.98%	
Target Compounds						
16) Acetone	3.81	43	17517	4.292	ug/l	88
18) Methyl Acetate	4.31	43	42119	1.517	ug/l	97
20) Methylene Chloride	4.54	84	21089	1.847	ug/l	98
43) Isopropyl Acetate	8.16	43	381491	18.385	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081919\
Data File : VN057400.D
Acq On : 19 Aug 2019 15:07
Operator : JC/SP
Sample : K4339-21
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T8-A4-(4)

Quant Time: Aug 20 08:18:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 11:33:14 2019
Response via : Initial Calibration

