

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081919\
 Data File : VN057404.D
 Acq On : 19 Aug 2019 16:45
 Operator : JC/SP
 Sample : K4272-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2S-K-(0)

Quant Time: Aug 20 08:27:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 11:33:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1001861	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1404336	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	1172525	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	309692	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	460236	42.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	84.52%	
35) Dibromofluoromethane	7.58	113	418065	45.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.82%	
50) Toluene-d8	10.09	98	1631185	48.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.90%	
62) 4-Bromofluorobenzene	12.40	95	428702	36.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	72.06%	
Target Compounds						
16) Acetone	3.81	43	29814	7.891	ug/l	98
18) Methyl Acetate	4.31	43	43427	1.978	ug/l	99
20) Methylene Chloride	4.55	84	21688	2.052	ug/l	91
43) Isopropyl Acetate	8.16	43	250652	12.653	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081919\
Data File : VN057404.D
Acq On : 19 Aug 2019 16:45
Operator : JC/SP
Sample : K4272-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2S-K-(0)

Quant Time: Aug 20 08:27:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 11:33:14 2019
Response via : Initial Calibration

