

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081919\
 Data File : VN057407.D
 Acq On : 19 Aug 2019 17:58
 Operator : JC/SP
 Sample : K4389-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T6-E3-(2)

Quant Time: Aug 20 08:36:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 11:33:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1063567	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1485661	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1234851	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	319684	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	493064	42.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.28%	
35) Dibromofluoromethane	7.58	113	451303	46.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.68%	
50) Toluene-d8	10.08	98	1709008	47.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.96%	
62) 4-Bromofluorobenzene	12.40	95	452852	35.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.96%	
Target Compounds						
16) Acetone	3.82	43	22342	5.570	ug/l	100
18) Methyl Acetate	4.31	43	36297	1.016	ug/l	96
43) Isopropyl Acetate	8.16	43	170197	7.708	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081919\
Data File : VN057407.D
Acq On : 19 Aug 2019 17:58
Operator : JC/SP
Sample : K4389-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T6-E3-(2)

Quant Time: Aug 20 08:36:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 11:33:14 2019
Response via : Initial Calibration

