

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081919\
 Data File : VN057410.D
 Acq On : 19 Aug 2019 19:10
 Operator : JC/SP
 Sample : K4389-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T4-F-(2)

Quant Time: Aug 20 08:42:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 11:33:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	987507	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1399085	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	1169742	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	312193	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	468519	43.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.28%	
35) Dibromofluoromethane	7.58	113	427122	46.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.14%	
50) Toluene-d8	10.09	98	1613341	48.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.20%	
62) 4-Bromofluorobenzene	12.40	95	423915	35.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.52%	
Target Compounds						
16) Acetone	3.81	43	20638	5.542	ug/l	91
20) Methylene Chloride	4.55	84	11212	1.076	ug/l	99
43) Isopropyl Acetate	8.16	43	263865	13.436	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081919\
Data File : VN057410.D
Acq On : 19 Aug 2019 19:10
Operator : JC/SP
Sample : K4389-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T4-F-(2)

Quant Time: Aug 20 08:42:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 11:33:14 2019
Response via : Initial Calibration

