

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081919\
 Data File : VN057415.D
 Acq On : 19 Aug 2019 22:24
 Operator : JC/SP
 Sample : VN0819WBL02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0819WBL02

Quant Time: Aug 20 08:59:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 11:33:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1049435	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1465776	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1207138	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	313999	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	484395	42.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	84.92%	
35) Dibromofluoromethane	7.58	113	446584	46.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.96%	
50) Toluene-d8	10.09	98	1668240	47.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.94%	
62) 4-Bromofluorobenzene	12.40	95	437442	35.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	70.44%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081919\
Data File : VN057415.D
Acq On : 19 Aug 2019 22:24
Operator : JC/SP
Sample : VN0819WBL02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0819WBL02

Quant Time: Aug 20 08:59:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 11:33:14 2019
Response via : Initial Calibration

