

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081919\
 Data File : VN057418.D
 Acq On : 19 Aug 2019 23:37
 Operator : JC/SP
 Sample : K4386-21
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T8-E2-(0)

Manual Integrations
 APPROVED

MMDadoda
 8/21/2019 9:45:41 AM

Quant Time: Aug 20 09:06:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 11:33:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	923689	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1271315	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	1055744	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	266758	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	415853	41.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	82.82%	
35) Dibromofluoromethane	7.58	113	382588	45.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.82%	
50) Toluene-d8	10.08	98	1462502	47.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.96%	
62) 4-Bromofluorobenzene	12.40	95	382099	35.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	70.94%	
Target Compounds						
16) Acetone	3.81	43	10507m	3.016	ug/l	
43) Isopropyl Acetate	8.16	43	162934	8.760	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081919\
Data File : VN057418.D
Acq On : 19 Aug 2019 23:37
Operator : JC/SP
Sample : K4386-21
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T8-E2-(0)

Manual Integrations
APPROVED

MMDadoda
8/21/2019 9:45:41 AM

Quant Time: Aug 20 09:06:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 11:33:14 2019
Response via : Initial Calibration

