

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081919\
 Data File : VN057419.D
 Acq On : 20 Aug 2019 00:01
 Operator : JC/SP
 Sample : K4386-22
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T8-E3-(0)

Quant Time: Aug 20 09:09:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 11:33:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1071807	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1468311	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1237849	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	322328	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	489357	42.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.00%	
35) Dibromofluoromethane	7.58	113	443600	46.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.18%	
50) Toluene-d8	10.08	98	1702916	48.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.74%	
62) 4-Bromofluorobenzene	12.40	95	449916	36.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.32%	
Target Compounds						
16) Acetone	3.82	43	13214	3.269	ug/l	95
43) Isopropyl Acetate	8.16	43	141879	6.321	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081919\
Data File : VN057419.D
Acq On : 20 Aug 2019 00:01
Operator : JC/SP
Sample : K4386-22
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T8-E3-(0)

Quant Time: Aug 20 09:09:23 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 11:33:14 2019
Response via : Initial Calibration

