

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081919\
 Data File : VN057425.D
 Acq On : 20 Aug 2019 2:26
 Operator : JC/SP
 Sample : K4386-29
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T8-F-(0)

Quant Time: Aug 20 09:25:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 11:33:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1020364	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1407653	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1143823	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	284778	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	455022	41.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.04%	
35) Dibromofluoromethane	7.58	113	424398	45.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.98%	
50) Toluene-d8	10.08	98	1616099	47.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.76%	
62) 4-Bromofluorobenzene	12.40	95	412592	34.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	69.18%	
Target Compounds						
16) Acetone	3.81	43	10701	2.781	ug/l	96
43) Isopropyl Acetate	8.16	43	177405	8.595	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081919\
Data File : VN057425.D
Acq On : 20 Aug 2019 2:26
Operator : JC/SP
Sample : K4386-29
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T8-F-(0)

Quant Time: Aug 20 09:25:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 11:33:14 2019
Response via : Initial Calibration

