

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081919\
 Data File : VN057432.D
 Acq On : 20 Aug 2019 5:16
 Operator : JC/SP
 Sample : K4386-37
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T6-A2-(2)

Quant Time: Aug 20 09:38:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 11:33:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1015769	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1403309	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1155726	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	294376	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	456189	41.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.62%	
35) Dibromofluoromethane	7.58	113	419078	45.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.12%	
50) Toluene-d8	10.08	98	1608851	47.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.64%	
62) 4-Bromofluorobenzene	12.40	95	412566	34.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	69.40%	
Target Compounds						
16) Acetone	3.82	43	11679	3.049	ug/l	95
43) Isopropyl Acetate	8.15	43	179615	8.747	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081919\
Data File : VN057432.D
Acq On : 20 Aug 2019 5:16
Operator : JC/SP
Sample : K4386-37
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T6-A2-(2)

Quant Time: Aug 20 09:38:09 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 11:33:14 2019
Response via : Initial Calibration

