

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081919\
 Data File : VN057434.D
 Acq On : 20 Aug 2019 6:04
 Operator : JC/SP
 Sample : K4386-39
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T6-A4-(2)

Quant Time: Aug 20 09:41:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 11:33:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1016176	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1417972	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1206469	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	332991	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	482469	43.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.34%	
35) Dibromofluoromethane	7.58	113	435194	46.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.64%	
50) Toluene-d8	10.08	98	1646215	48.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.84%	
62) 4-Bromofluorobenzene	12.40	95	445740	37.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.20%	
Target Compounds						
16) Acetone	3.81	43	12214	3.187	ug/l	95
43) Isopropyl Acetate	8.16	43	159817	7.565	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081919\
Data File : VN057434.D
Acq On : 20 Aug 2019 6:04
Operator : JC/SP
Sample : K4386-39
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T6-A4-(2)

Quant Time: Aug 20 09:41:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081519W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 11:33:14 2019
Response via : Initial Calibration

