

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081920\
 Data File : VN063006.D
 Acq On : 19 Aug 2020 12:13
 Operator : JC/MD
 Sample : L3718-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB181-TB-20200812

Quant Time: Aug 19 14:18:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 19 06:09:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	149822	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	249832	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	254797	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	94819	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	98368	55.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.80%	
35) Dibromofluoromethane	7.55	113	86201	52.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.90%	
50) Toluene-d8	10.06	98	320009	51.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.36%	
62) 4-Bromofluorobenzene	12.38	95	106479	48.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.52%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081920\
Data File : VN063006.D
Acq On : 19 Aug 2020 12:13
Operator : JC/MD
Sample : L3718-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB181-TB-20200812

Quant Time: Aug 19 14:18:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081820W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 19 06:09:27 2020
Response via : Initial Calibration

