

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081920\
 Data File : VN063010.D
 Acq On : 19 Aug 2020 13:50
 Operator : JC/MD
 Sample : L3718-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB181-GW-218-220

Quant Time: Aug 19 15:34:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 19 06:09:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	144164	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	234380	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	245060	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	91592	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	89933	52.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.28%	
35) Dibromofluoromethane	7.55	113	80781	52.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.78%	
50) Toluene-d8	10.06	98	305842	52.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.28%	
62) 4-Bromofluorobenzene	12.38	95	101223	48.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.80%	
Target Compounds						
30) Chloroform	7.33	83	2373	0.861	ug/l	94
44) Trichloroethene	8.80	130	7033	4.316	ug/l	93
64) Tetrachloroethene	10.60	164	1161	0.721	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081920\
Data File : VN063010.D
Acq On : 19 Aug 2020 13:50
Operator : JC/MD
Sample : L3718-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB181-GW-218-220

Quant Time: Aug 19 15:34:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081820W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 19 06:09:27 2020
Response via : Initial Calibration

