

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081920\
 Data File : VN063012.D
 Acq On : 19 Aug 2020 14:39
 Operator : JC/MD
 Sample : L3718-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB181-GW-238-240

Quant Time: Aug 19 15:40:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 19 06:09:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	148859	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	247881	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	258898	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	94446	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	94260	53.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.86%	
35) Dibromofluoromethane	7.55	113	84404	52.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.52%	
50) Toluene-d8	10.06	98	320075	51.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.18%	
62) 4-Bromofluorobenzene	12.38	95	108304	49.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.94%	
Target Compounds						
27) cis-1,2-Dichloroethene	6.78	96	16188	9.767	ug/l	91
30) Chloroform	7.33	83	1638	0.576	ug/l	99
44) Trichloroethene	8.81	130	4010	2.327	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081920\
Data File : VN063012.D
Acq On : 19 Aug 2020 14:39
Operator : JC/MD
Sample : L3718-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB181-GW-238-240

Quant Time: Aug 19 15:40:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081820W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 19 06:09:27 2020
Response via : Initial Calibration

