

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081920\
 Data File : VN063017.D
 Acq On : 19 Aug 2020 17:09
 Operator : JC/MD
 Sample : PB131068TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131068TB

Quant Time: Aug 20 01:40:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 19 06:09:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	144779	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	247021	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	267595	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	95546	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	96131	56.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.04%	
35) Dibromofluoromethane	7.55	113	83838	52.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.18%	
50) Toluene-d8	10.06	98	329193	53.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.50%	
62) 4-Bromofluorobenzene	12.38	95	112610	51.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.24%	
Target Compounds						
16) Acetone	3.78	43	9156	13.399	ug/l	100
18) Methyl Acetate	4.30	43	2461	1.443	ug/l #	63
20) Methylene Chloride	4.51	84	5419	2.308	ug/l	99
43) Isopropyl Acetate	8.13	43	40165	11.886	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081920\
Data File : VN063017.D
Acq On : 19 Aug 2020 17:09
Operator : JC/MD
Sample : PB131068TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131068TB

Quant Time: Aug 20 01:40:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081820W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 19 06:09:27 2020
Response via : Initial Calibration

