

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081920\
 Data File : VN063018.D
 Acq On : 19 Aug 2020 17:33
 Operator : JC/MD
 Sample : L3496-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AU-06-081720

Quant Time: Aug 20 01:44:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 19 06:09:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	147771	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	246733	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	252714	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	92242	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	96644	55.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.36%	
35) Dibromofluoromethane	7.55	113	85164	52.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.94%	
50) Toluene-d8	10.06	98	316055	51.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.36%	
62) 4-Bromofluorobenzene	12.38	95	105122	48.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.48%	
Target Compounds						
20) Methylene Chloride	4.50	84	8470	4.264	ug/l	96
30) Chloroform	7.33	83	3212	1.137	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081920\
Data File : VN063018.D
Acq On : 19 Aug 2020 17:33
Operator : JC/MD
Sample : L3496-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AU-06-081720

Quant Time: Aug 20 01:44:48 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081820W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 19 06:09:27 2020
Response via : Initial Calibration

