

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081920\
 Data File : VN063022.D
 Acq On : 19 Aug 2020 19:10
 Operator : JC/MD
 Sample : L3736-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OUTFALL001-200818

Quant Time: Aug 20 01:57:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 19 06:09:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	142321	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	244761	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	246919	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	89961	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	90860	53.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.74%	
35) Dibromofluoromethane	7.55	113	81353	51.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.02%	
50) Toluene-d8	10.06	98	310871	50.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.50%	
62) 4-Bromofluorobenzene	12.38	95	102406	47.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.74%	
Target Compounds						
27) cis-1,2-Dichloroethene	6.78	96	2133	1.346	ug/l	Qvalue 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081920\
Data File : VN063022.D
Acq On : 19 Aug 2020 19:10
Operator : JC/MD
Sample : L3736-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OUTFALL001-200818

Quant Time: Aug 20 01:57:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081820W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 19 06:09:27 2020
Response via : Initial Calibration

