

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081920\
 Data File : VN063023.D
 Acq On : 19 Aug 2020 19:35
 Operator : JC/MD
 Sample : L3736-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-02-200818

Quant Time: Aug 20 02:00:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 19 06:09:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	160433	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	269941	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	276830	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	102670	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	100039	52.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.22%	
35) Dibromofluoromethane	7.55	113	90131	51.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.48%	
50) Toluene-d8	10.06	98	343812	50.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.78%	
62) 4-Bromofluorobenzene	12.38	95	114778	48.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.28%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081920\
Data File : VN063023.D
Acq On : 19 Aug 2020 19:35
Operator : JC/MD
Sample : L3736-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-02-200818

Quant Time: Aug 20 02:00:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081820W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 19 06:09:27 2020
Response via : Initial Calibration

