

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081921\
 Data File : VN068210.D
 Acq On : 19 Aug 2021 12:15
 Operator : JC/MD
 Sample : VN0819WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0819WBL01

Quant Time: Aug 19 13:03:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	446785	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	746295	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	739891	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	276994	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	278825	58.448	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 116.900%	
35) Dibromofluoromethane	8.024	113	254956	54.415	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 108.820%	
50) Toluene-d8	10.446	98	986094	55.234	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 110.460%	
62) 4-Bromofluorobenzene	12.736	95	303167	51.010	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 102.020%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081921\
Data File : VN068210.D
Acq On : 19 Aug 2021 12:15
Operator : JC/MD
Sample : VN0819WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0819WBL01

Quant Time: Aug 19 13:03:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

