

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081921\
 Data File : VN068214.D
 Acq On : 19 Aug 2021 14:10
 Operator : JC/MD
 Sample : M3410-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB

Quant Time: Aug 19 14:27:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	469886	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	780727	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	767333	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	285175	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	282397	56.287	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	112.580%
35) Dibromofluoromethane	8.024	113	246998	50.391	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.780%
50) Toluene-d8	10.443	98	974576	52.181	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.360%
62) 4-Bromofluorobenzene	12.736	95	304841	49.030	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081921\
Data File : VN068214.D
Acq On : 19 Aug 2021 14:10
Operator : JC/MD
Sample : M3410-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB

Quant Time: Aug 19 14:27:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

