

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081921\
 Data File : VN068216.D
 Acq On : 19 Aug 2021 14:59
 Operator : JC/MD
 Sample : M3463-45
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 2021-08-17-TB

Manual Integrations
 APPROVED

MMDadoda
 8/20/2021 6:50:16 PM

Quant Time: Aug 20 01:54:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	446442	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	752838	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	755811	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	280232	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	271749	57.009	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	= 114.020%		
35) Dibromofluoromethane	8.024	113	243905	51.604	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	= 103.200%		
50) Toluene-d8	10.446	98	965604	53.616	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	= 107.240%		
62) 4-Bromofluorobenzene	12.733	95	308265	51.417	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	= 102.840%		
Target Compounds						Qvalue
16) Acetone	4.296	43	4098m	2.944	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081921\
Data File : VN068216.D
Acq On : 19 Aug 2021 14:59
Operator : JC/MD
Sample : M3463-45
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
2021-08-17-TB

Manual Integrations
APPROVED

MMDadoda
8/20/2021 6:50:16 PM

Quant Time: Aug 20 01:54:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

