

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081921\
 Data File : VN068219.D
 Acq On : 19 Aug 2021 16:14
 Operator : JC/MD
 Sample : M3449-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP2

Quant Time: Aug 20 01:54:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	470655	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	801840	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	877040	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	322496	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	269683	53.665	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 107.320%	
35) Dibromofluoromethane	8.024	113	244470	48.562	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 97.120%	
50) Toluene-d8	10.443	98	1031272	53.763	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 107.520%	
62) 4-Bromofluorobenzene	12.734	95	329946	51.671	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 103.340%	
Target Compounds						Qvalue
16) Acetone	4.293	43	12249	8.348	ug/l	91
20) Methylene Chloride	5.093	84	34450	6.643	ug/l #	78
43) Isopropyl Acetate	8.566	43	35427	3.961	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081921\
Data File : VN068219.D
Acq On : 19 Aug 2021 16:14
Operator : JC/MD
Sample : M3449-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP2

Quant Time: Aug 20 01:54:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

