

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081921\
 Data File : VN068220.D
 Acq On : 19 Aug 2021 16:39
 Operator : JC/MD
 Sample : M3449-08
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP3

Quant Time: Aug 20 01:54:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	454333	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	778096	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	811754	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	305377	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	271579	55.984	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.960%
35) Dibromofluoromethane	8.024	113	243844	49.916	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.840%
50) Toluene-d8	10.446	98	989685	53.169	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.340%
62) 4-Bromofluorobenzene	12.734	95	314022	50.677	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.360%
Target Compounds						
						Qvalue
16) Acetone	4.299	43	11880	8.388	ug/l	95
20) Methylene Chloride	5.098	84	32887	6.545	ug/l #	82
43) Isopropyl Acetate	8.566	43	35590	4.100	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081921\
Data File : VN068220.D
Acq On : 19 Aug 2021 16:39
Operator : JC/MD
Sample : M3449-08
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP3

Quant Time: Aug 20 01:54:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

