

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081921\
 Data File : VN068221.D
 Acq On : 19 Aug 2021 17:04
 Operator : JC/MD
 Sample : PB138556TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138556TB

Quant Time: Aug 20 01:55:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	463083	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	812943	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	853716	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	317338	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	267993	54.201	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	108.400%
35) Dibromofluoromethane	8.024	113	238183	46.667	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.340%
50) Toluene-d8	10.443	98	979498	50.366	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.740%
62) 4-Bromofluorobenzene	12.736	95	321152	49.606	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.220%
Target Compounds						
						Qvalue
16) Acetone	4.296	43	26366	18.264	ug/l	92
20) Methylene Chloride	5.095	84	24802	4.265	ug/l	# 80
43) Isopropyl Acetate	8.566	43	38652	4.262	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081921\
Data File : VN068221.D
Acq On : 19 Aug 2021 17:04
Operator : JC/MD
Sample : PB138556TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB138556TB

Quant Time: Aug 20 01:55:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

