

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081921\
 Data File : VN068225.D
 Acq On : 19 Aug 2021 18:43
 Operator : JC/MD
 Sample : PB138556ZHE#04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138556ZHE#04

Quant Time: Aug 20 01:55:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	412698	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	725048	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	771753	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	286629	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	239662	54.388	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	= 108.780%		
35) Dibromofluoromethane	8.024	113	217813	47.849	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	= 95.700%		
50) Toluene-d8	10.446	98	894272	51.558	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	= 103.120%		
62) 4-Bromofluorobenzene	12.736	95	288283	49.928	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	= 99.860%		
Target Compounds						Qvalue
16) Acetone	4.296	43	10195	7.924	ug/l	99
20) Methylene Chloride	5.092	84	30809	6.819	ug/l #	76
43) Isopropyl Acetate	8.566	43	42673	5.276	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081921\
Data File : VN068225.D
Acq On : 19 Aug 2021 18:43
Operator : JC/MD
Sample : PB138556ZHE#04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB138556ZHE#04

Quant Time: Aug 20 01:55:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

