

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081921\
 Data File : VN068228.D
 Acq On : 19 Aug 2021 19:58
 Operator : JC/MD
 Sample : PB138556ZHE#07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138556ZHE#07

Quant Time: Aug 20 01:56:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	401802	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	695787	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	729403	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	271060	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	236568	55.142	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	=	110.280%	
35) Dibromofluoromethane	8.024	113	214528	49.110	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	=	98.220%	
50) Toluene-d8	10.446	98	860138	51.676	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	=	103.360%	
62) 4-Bromofluorobenzene	12.733	95	275693	49.755	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	=	99.520%	
Target Compounds						Qvalue
16) Acetone	4.293	43	12328	9.842	ug/l	# 85
20) Methylene Chloride	5.103	84	33773	7.958	ug/l	# 80
43) Isopropyl Acetate	8.563	43	40253	5.186	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081921\
Data File : VN068228.D
Acq On : 19 Aug 2021 19:58
Operator : JC/MD
Sample : PB138556ZHE#07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB138556ZHE#07

Quant Time: Aug 20 01:56:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

