

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081921\
 Data File : VN068230.D
 Acq On : 19 Aug 2021 20:47
 Operator : JC/MD
 Sample : PB138556ZHE#09
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138556ZHE#09

Quant Time: Aug 20 01:56:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	405442	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	705894	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	755641	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.680	152	275728	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	238388	55.068	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	110.140%
35) Dibromofluoromethane	8.027	113	212810	48.019	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.040%
50) Toluene-d8	10.446	98	869849	51.511	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.020%
62) 4-Bromofluorobenzene	12.736	95	277057	49.285	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.580%
Target Compounds						
						Qvalue
16) Acetone	4.296	43	24439	19.336	ug/l	91
20) Methylene Chloride	5.098	84	21076	4.074	ug/l #	79
43) Isopropyl Acetate	8.568	43	41026	5.210	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081921\
Data File : VN068230.D
Acq On : 19 Aug 2021 20:47
Operator : JC/MD
Sample : PB138556ZHE#09
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB138556ZHE#09

Quant Time: Aug 20 01:56:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

