

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081921\
 Data File : VN068231.D
 Acq On : 19 Aug 2021 21:12
 Operator : JC/MD
 Sample : PB138556ZHE#10
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138556ZHE#10

Quant Time: Aug 20 01:56:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	431152	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.968	114	750769	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	787173	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	285974	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	250699	54.458	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	108.920%
35) Dibromofluoromethane	8.024	113	227176	48.197	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.400%
50) Toluene-d8	10.443	98	922557	51.367	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.740%
62) 4-Bromofluorobenzene	12.734	95	290980	48.668	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.340%
Target Compounds						
						Qvalue
16) Acetone	4.299	43	27135	20.188	ug/l	94
20) Methylene Chloride	5.098	84	21340	3.773	ug/l #	81
43) Isopropyl Acetate	8.568	43	39621	4.731	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081921\
Data File : VN068231.D
Acq On : 19 Aug 2021 21:12
Operator : JC/MD
Sample : PB138556ZHE#10
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB138556ZHE#10

Quant Time: Aug 20 01:56:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

