

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081921\
 Data File : VN068232.D
 Acq On : 19 Aug 2021 21:37
 Operator : JC/MD
 Sample : PB138556ZHE#11
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138556ZHE#11

Quant Time: Aug 20 01:56:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 13:09:49 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	414220	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	724982	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	772809	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	291775	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	241375	54.576	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 109.160%	
35) Dibromofluoromethane	8.024	113	216298	47.521	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 95.040%	
50) Toluene-d8	10.446	98	908440	52.380	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 104.760%	
62) 4-Bromofluorobenzene	12.736	95	288180	49.914	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 99.820%	
Target Compounds						
						Qvalue
16) Acetone	4.296	43	26205	20.293	ug/l	# 88
20) Methylene Chloride	5.092	84	23150	4.547	ug/l	# 77
43) Isopropyl Acetate	8.566	43	37685	4.660	ug/l	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081921\
Data File : VN068232.D
Acq On : 19 Aug 2021 21:37
Operator : JC/MD
Sample : PB138556ZHE#11
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB138556ZHE#11

Quant Time: Aug 20 01:56:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 13:09:49 2021
Response via : Initial Calibration

