

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081922\
 Data File : VN074025.D
 Acq On : 19 Aug 2022 14:35
 Operator : JC\MD
 Sample : N4264-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 U-3

Quant Time: Aug 20 02:46:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N072722W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Jul 27 18:07:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Bromochloromethane	7.586	128	67304	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.904	114	341691	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.686	117	295717	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.375	65	168453	28.948	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	96.500%
60) 4-Bromofluorobenzene	12.674	95	138294	26.412	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	88.033%
63) Toluene-d8	10.380	98	481300	28.905	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	96.333%
Target Compounds						
					Qvalue	
15) Acetone	4.222	58	1266	1.378	ug/l #	19
25) Chloroform	7.751	83	47374	5.087	ug/l	96
53) Dibromochloromethane	10.916	129	10791	2.007	ug/l #	10
56) Bromoform	12.668	173	1042	0.247	ug/l #	1
61) Tetrachloroethene	10.916	164	12595	2.781	ug/l #	83
77) t-1,4-Dichloro-2-butene	12.668	75	66255	29.013	ug/l #	8
91) Naphthalene	15.439	128	4377	0.247	ug/l #	81


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081922\
Data File : VN074025.D
Acq On : 19 Aug 2022 14:35
Operator : JC\MD
Sample : N4264-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
U-3

Quant Time: Aug 20 02:46:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N072722W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Jul 27 18:07:26 2022
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 30.000 ug/l
RT: 7.586 min Scan# 969
Delta R.T. 0.000 min
Lab File: VN074025.D
Acq: 19 Aug 2022 14:35




Instrument :
MSVOA_N
ClientSampleId :
U-3


Tgt Ion:128 Resp: 67304
Ion Ratio Lower Upper
128 100
49 128.9 0.0 392.3
130 127.9 0.0 320.2



#15
Acetone
Concen: 1.378 ug/l
RT: 4.222 min Scan# 397
Delta R.T. -0.006 min
Lab File: VN074025.D
Acq: 19 Aug 2022 14:35

Tgt Ion: 58 Resp: 1266
Ion Ratio Lower Upper
58 100
43 154.2 256.6 384.8#









#25
Chloroform
Concen: 5.087 ug/l
RT: 7.751 min Scan# 996
Delta R.T. 0.006 min
Lab File: VN074025.D
Acq: 19 Aug 2022 14:35




Instrument :
MSVOA_N
ClientSampleId :
U-3


Tgt Ion: 83 Resp: 47374
Ion Ratio Lower Upper
83 100
85 60.8 51.3 76.9



#27
1,2-Dichloroethane-d4
Concen: 28.948 ug/l
RT: 8.375 min Scan# 1103
Delta R.T. 0.006 min
Lab File: VN074025.D
Acq: 19 Aug 2022 14:35

Tgt Ion: 65 Resp: 168453
Ion Ratio Lower Upper
65 100
67 51.8 40.2 60.2









#28
1,4-Difluorobenzene
Concen: 30.000 ug/l
RT: 8.904 min Scan# 1193
Delta R.T. -0.000 min
Lab File: VN074025.D
Acq: 19 Aug 2022 14:35




Instrument :
MSVOA_N
ClientSampleId :
U-3


Tgt Ion:114 Resp: 341691
Ion Ratio Lower Upper
114 100
63 18.4 15.8 23.6
88 14.9 11.6 17.4



#53
Dibromochloromethane
Concen: 2.007 ug/l
RT: 10.916 min Scan# 1535
Delta R.T. -0.259 min
Lab File: VN074025.D
Acq: 19 Aug 2022 14:35

Tgt Ion:129 Resp: 10791
Ion Ratio Lower Upper
129 100
127 0.0 39.0 117.0#









#56
Bromoform
Concen: 0.247 ug/l
RT: 12.668 min Scan# 1833
Delta R.T. 0.270 min
Lab File: VN074025.D
Acq: 19 Aug 2022 14:35




Instrument :
MSVOA_N
ClientSampleId :
U-3


Tgt Ion:173 Resp: 1042
Ion Ratio Lower Upper
173 100
175 894.5 38.6 58.0#
254 0.0 0.0 0.0



#57
Chlorobenzene-d5
Concen: 30.000 ug/l
RT: 11.686 min Scan# 1666
Delta R.T. -0.000 min
Lab File: VN074025.D
Acq: 19 Aug 2022 14:35

Tgt Ion:117 Resp: 295717
Ion Ratio Lower Upper
117 100
82 51.0 43.6 65.4
119 32.1 25.8 38.6









#60
4-Bromofluorobenzene
Concen: 26.412 ug/l
RT: 12.674 min Scan# 1834
Delta R.T. -0.000 min
Lab File: VN074025.D
Acq: 19 Aug 2022 14:35




Instrument :
MSVOA_N
ClientSampleId :
U-3


Tgt Ion: 95 Resp: 138294
Ion Ratio Lower Upper
95 100
174 93.1 62.9 94.3
176 88.9 61.0 91.4



#61
Tetrachloroethene
Concen: 2.781 ug/l
RT: 10.916 min Scan# 1535
Delta R.T. -0.000 min
Lab File: VN074025.D
Acq: 19 Aug 2022 14:35

Tgt Ion: 164 Resp: 12595
Ion Ratio Lower Upper
164 100
166 126.2 105.8 158.8
129 77.2 82.5 123.7#
131 76.7 81.2 121.8#









#63
Toluene-d8
Concen: 28.905 ug/l
RT: 10.380 min Scan# 1444
Delta R.T. -0.000 min
Lab File: VN074025.D
Acq: 19 Aug 2022 14:35




Instrument :
MSVOA_N
ClientSampleId :
U-3


Tgt Ion: 98 Resp: 481300
Ion Ratio Lower Upper
98 100
100 63.7 50.6 76.0



#77
t-1,4-Dichloro-2-butene
Concen: 29.013 ug/l
RT: 12.668 min Scan# 1833
Delta R.T. 0.100 min
Lab File: VN074025.D
Acq: 19 Aug 2022 14:35


Tgt Ion: 75 Resp: 66255
Ion Ratio Lower Upper
75 100
53 0.0 51.8 155.5#
89 0.0 22.9 68.7#





#91
 Naphthalene
 Concen: 0.247 ug/l
 RT: 15.439 min Scan# 21
 Delta R.T. 0.006 min
 Lab File: VN074025.D
 Acq: 19 Aug 2022 14:35

Instrument :
 MSVOA_N
 ClientSampleId :
 U-3



Tgt Ion:	128	Resp:	4377
Ion	Ratio	Lower	Upper
128	100		
127	8.4	10.7	16.1#
129	1.3	9.0	13.6#

