

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081922\
 Data File : VN074025.D
 Acq On : 19 Aug 2022 14:35
 Operator : JC\MD
 Sample : N4264-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 U-3

Quant Time: Aug 20 02:46:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N072722W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Jul 27 18:07:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.586	128	67304	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.904	114	341691	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.686	117	295717	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.375	65	168453	28.948	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	96.500%
60) 4-Bromofluorobenzene	12.674	95	138294	26.412	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	88.033%
63) Toluene-d8	10.380	98	481300	28.905	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	96.333%
Target Compounds						
25) Chloroform	7.751	83	47374	5.087	ug/l	96
61) Tetrachloroethene	10.916	164	12595	2.781	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081922\
Data File : VN074025.D
Acq On : 19 Aug 2022 14:35
Operator : JC\MD
Sample : N4264-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
U-3

Quant Time: Aug 20 02:46:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N072722W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Jul 27 18:07:26 2022
Response via : Initial Calibration

