

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081922\
 Data File : VN074031.D
 Acq On : 19 Aug 2022 17:02
 Operator : JC\MD
 Sample : PB147119ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB147119ZHE#02

Quant Time: Aug 20 03:05:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	547779	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	853641	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	856598	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	453751	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	285602	53.530	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.060%			
35) Dibromofluoromethane	7.951	113	273311	54.139	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 108.280%			
50) Toluene-d8	10.381	98	927675	47.447	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 94.900%			
62) 4-Bromofluorobenzene	12.674	95	393940	60.710	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 121.420%			
Target Compounds						
16) Acetone	4.228	43	76868	30.538	ug/l	99
20) Methylene Chloride	5.010	84	35572	6.697	ug/l	94
37) Ethyl Acetate	7.340	43	77686	9.145	ug/l	99
43) Isopropyl Acetate	8.498	43	318785	26.342	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081922\
Data File : VN074031.D
Acq On : 19 Aug 2022 17:02
Operator : JC\MD
Sample : PB147119ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB147119ZHE#02

Quant Time: Aug 20 03:05:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

