

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081922\
 Data File : VN074034.D
 Acq On : 19 Aug 2022 18:17
 Operator : JC\MD
 Sample : PB147119ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB147119ZHE#05

Quant Time: Aug 20 03:07:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	523022	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	832097	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	832654	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	467079	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	279569	54.880	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.760%	
35) Dibromofluoromethane	7.951	113	268262	54.514	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.020%	
50) Toluene-d8	10.381	98	900347	47.242	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 94.480%	
62) 4-Bromofluorobenzene	12.675	95	388970	61.496	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 123.000%	
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	75326	31.342	ug/l	95
20) Methylene Chloride	5.022	84	34809	6.864	ug/l	96
37) Ethyl Acetate	7.340	43	74370	8.981	ug/l	98
43) Isopropyl Acetate	8.498	43	327280	27.744	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081922\
Data File : VN074034.D
Acq On : 19 Aug 2022 18:17
Operator : JC\MD
Sample : PB147119ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB147119ZHE#05

Quant Time: Aug 20 03:07:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

