

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081922\
 Data File : VN074035.D
 Acq On : 19 Aug 2022 18:41
 Operator : JC\MD
 Sample : PB147119ZHE#06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB147119ZHE#06

Quant Time: Aug 20 03:07:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	529998	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	857183	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	864428	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	485699	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	284432	55.099	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 110.200%		
35) Dibromofluoromethane	7.951	113	273576	53.967	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 107.940%		
50) Toluene-d8	10.380	98	923900	47.059	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 94.120%		
62) 4-Bromofluorobenzene	12.674	95	407373	62.521	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 125.040%#		
Target Compounds						
16) Acetone	4.228	43	75958	31.189	ug/l	98
20) Methylene Chloride	5.016	84	36689	7.139	ug/l	98
37) Ethyl Acetate	7.339	43	75008	8.793	ug/l	97
43) Isopropyl Acetate	8.498	43	326772	26.890	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081922\
Data File : VN074035.D
Acq On : 19 Aug 2022 18:41
Operator : JC\MD
Sample : PB147119ZHE#06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB147119ZHE#06

Quant Time: Aug 20 03:07:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

