

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081922\
 Data File : VN074040.D
 Acq On : 19 Aug 2022 20:45
 Operator : JC\MD
 Sample : N4247-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-3

Quant Time: Aug 20 03:09:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	514318	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	830546	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	838408	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	484939	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	276527	55.201	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.400%
35) Dibromofluoromethane	7.957	113	261302	53.199	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.400%
50) Toluene-d8	10.380	98	903856	47.514	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.020%
62) 4-Bromofluorobenzene	12.674	95	399635	63.300	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	126.600%#
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	59949	25.366	ug/l	98
20) Methylene Chloride	5.016	84	35830	7.185	ug/l	97
37) Ethyl Acetate	7.339	43	77143	9.334	ug/l	97
43) Isopropyl Acetate	8.498	43	332290	28.222	ug/l #	91
95) Naphthalene	15.433	128	59332	1.747	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081922\
Data File : VN074040.D
Acq On : 19 Aug 2022 20:45
Operator : JC\MD
Sample : N4247-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-3

Quant Time: Aug 20 03:09:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

