

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081922\
 Data File : VN074041.D
 Acq On : 19 Aug 2022 21:10
 Operator : JC\MD
 Sample : N4247-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-4

Quant Time: Aug 20 03:09:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 16 00:45:21 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	515674	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	821462	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	845121	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	505065	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	279309	55.610	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.220%
35) Dibromofluoromethane	7.957	113	263660	54.273	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.540%
50) Toluene-d8	10.380	98	906301	48.170	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.340%
62) 4-Bromofluorobenzene	12.674	95	406664	65.126	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	130.260%#
Target Compounds						
						Qvalue
16) Acetone	4.228	43	62916	26.551	ug/l	98
20) Methylene Chloride	5.010	84	34321	6.864	ug/l	91
37) Ethyl Acetate	7.345	43	79952	9.780	ug/l	99
43) Isopropyl Acetate	8.498	43	339410	29.145	ug/l #	90
52) Toluene	10.445	92	16628	1.139	ug/l	94
95) Naphthalene	15.433	128	62549	1.768	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081922\
Data File : VN074041.D
Acq On : 19 Aug 2022 21:10
Operator : JC\MD
Sample : N4247-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-4

Quant Time: Aug 20 03:09:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 16 00:45:21 2022
Response via : Initial Calibration

