


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081922\
 Data File : VN074017.D
 Acq On : 19 Aug 2022 10:25
 Operator : JC\MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N081522W.M
 Title : SW846 8260
 Last Update : Tue Aug 16 00:45:21 2022



AutoFind: Scans 1832, 1833, 1834; Background Corrected with Scan 1824

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.4	16682	PASS
75	95	30	60	50.8	48597	PASS
95	95	100	100	100.0	95627	PASS
96	95	5	9	6.8	6493	PASS
173	174	0.00	2	0.8	705	PASS
174	95	50	100	90.8	86808	PASS
175	174	5	9	7.6	6639	PASS
176	174	95	101	98.3	85355	PASS
177	176	5	9	6.7	5714	PASS