

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082018\
 Data File : VN050771.D
 Acq On : 20 Aug 2018 13:56
 Operator : MD\SY
 Sample : J4555-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TWP-02

Manual Integrations
 APPROVED

MMDadoda
 8/21/2018 2:13:19 PM

Quant Time: Aug 21 14:50:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N072518W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jul 26 15:21:50 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	107942	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	547609	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	465497	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	213599	29.11	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	97.03%
60) 4-Bromofluorobenzene	12.40	95	153984	21.53	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	71.77%
63) Toluene-d8	10.09	98	648480	29.17	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	97.23%

Target Compounds

					Ovalue
15) Acetone	3.82	58	8806	15.06 ug/l #	64
62) Toluene	10.16	91	191333	7.20 ug/l	98
91) Naphthalene	15.14	128	5558	7.95 ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN082018\
Data File : VN050771.D
Acq On : 20 Aug 2018 13:56
Operator : MD\SY
Sample : J4555-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TWP-02

Manual Integrations
APPROVED

MMDadoda
8/21/2018 2:13:19 PM

Quant Time: Aug 21 14:50:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N072518W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Jul 26 15:21:50 2018
Response via : Initial Calibration

