

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082018\
 Data File : VN050774.D
 Acq On : 20 Aug 2018 15:10
 Operator : MD\SY
 Sample : PB112183ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112183ZHE#01

Quant Time: Aug 21 04:02:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	648374	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1072124	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	927039	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	326606	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	431319	52.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.54%	
35) Dibromofluoromethane	7.59	113	406527	47.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.98%	
50) Toluene-d8	10.09	98	1475287	45.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.60%	
62) 4-Bromofluorobenzene	12.40	95	407345	38.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.56%	
Target Compounds						
16) Acetone	3.82	43	16660	6.75	ug/l	Qvalue 100
18) Methyl Acetate	4.34	43	11544	1.04	ug/l #	77
20) Methylene Chloride	4.55	84	36318	3.49	ug/l	89
43) Isopropyl Acetate	8.17	43	274692	22.34	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN082018\
Data File : VN050774.D
Acq On : 20 Aug 2018 15:10
Operator : MD\SY
Sample : PB112183ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112183ZHE#01

Quant Time: Aug 21 04:02:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

