

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082020\
 Data File : VN063041.D
 Acq On : 20 Aug 2020 17:12
 Operator : JC/MD
 Sample : L3741-02DL 20X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DSN-002-SEWER-OUTLET-002DL

Quant Time: Aug 21 06:28:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N082020W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Aug 20 14:01:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.15	128	34883	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.55	114	164116	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.38	117	152979	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	7.99	65	75340	30.48	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	101.60%
60) 4-Bromofluorobenzene	12.38	95	56195	23.25	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	77.50%
63) Toluene-d8	10.06	98	207232	29.52	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	98.40%

Target Compounds

					Ovalue
15) Acetone	3.78	58	16117	58.613	ug/l 98
18) Methylene Chloride	4.51	84	6774	2.974	ug/l 96
25) Chloroform	7.33	83	2264	0.592	ug/l 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082020\
Data File : VN063041.D
Acq On : 20 Aug 2020 17:12
Operator : JC/MD
Sample : L3741-02DL 20X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DSN-002-SEWER-OUTLET-002DL

Quant Time: Aug 21 06:28:32 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N082020W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Aug 20 14:01:17 2020
Response via : Initial Calibration

