

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082020\
 Data File : VN063043.D
 Acq On : 20 Aug 2020 18:01
 Operator : JC/MD
 Sample : L3723-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MANHOLE

Quant Time: Aug 21 06:33:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N082020W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Aug 20 14:01:17 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.15	128	30244	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.55	114	144566	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.38	117	134652	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	7.99	65	66720	31.14	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	103.80%
60) 4-Bromofluorobenzene	12.38	95	50759	23.86	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	79.53%
63) Toluene-d8	10.06	98	179005	28.97	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	96.57%

Target Compounds

					Ovalue
15) Acetone	3.78	58	13989	58.678	ug/l 98
25) Chloroform	7.33	83	19592	5.910	ug/l 97
41) Bromodichloromethane	9.37	83	7954	3.157	ug/l 94
53) Dibromochloromethane	10.88	129	3578	1.860	ug/l 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN082020\
Data File : VN063043.D
Acq On : 20 Aug 2020 18:01
Operator : JC/MD
Sample : L3723-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MANHOLE

Quant Time: Aug 21 06:33:47 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N082020W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Aug 20 14:01:17 2020
Response via : Initial Calibration

