

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN082020\
 Data File : VN063047.D
 Acq On : 20 Aug 2020 20:50
 Operator : JC/MD
 Sample : VN0820WBL02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0820WBL02

Quant Time: Aug 21 06:49:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 19 06:09:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	142978	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	238750	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	242888	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	89555	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	96455	56.92	ug/l	0.00
Spiked Amount			Recovery	=	113.84%	
35) Dibromofluoromethane	7.55	113	82548	53.06	ug/l	0.00
Spiked Amount			Recovery	=	106.12%	
50) Toluene-d8	10.06	98	305798	51.18	ug/l	0.00
Spiked Amount			Recovery	=	102.36%	
62) 4-Bromofluorobenzene	12.38	95	101072	47.93	ug/l	0.00
Spiked Amount			Recovery	=	95.86%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN082020\
Data File : VN063047.D
Acq On : 20 Aug 2020 20:50
Operator : JC/MD
Sample : VN0820WBL02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0820WBL02

Quant Time: Aug 21 06:49:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081820W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 19 06:09:27 2020
Response via : Initial Calibration

